3. *510(k) Summary of Safety and Effectiveness*

a. General Information

Modified Device Information

Category:	**Comments:**
Sponsor:	Boston Scientific Corporation 2710 Orchard Parkway San Jose, Ca 95134
Correspondent:	April I. Malmborg Senior Specialist, Regulatory Affairs Boston Scientific Corporation 2710 Orchard Parkway San Jose, Ca 95134
Contact Information:	E-mail: malmbora@bsci.com Phone: (408) 895-3637 Fax: (408) 895-2202
Device Common Name:	Electrosurgical Probe
Device Proprietary Name:	Cobra Surgical Probe
Device Classification:	21 CFR §878.4400

b. Predicate Device Information

Predicate Device:	Cobra Surgical Probe (K013873)
Predicate Device Manufacturer:	Boston Scientific Corporation
Predicate Device Common Name	Electrosurgical Probe
Predicate Device Classification:	21 CFR §878.4400
Predicate Device Classification Number:	Class II

c. Date Summary Prepared

January 30, 2004

d. Description of Device

The Cobra Surgical Probes are Electrosurgical Probes, with either a malleable or flexible shaft, used in conjunction with the Cobra Electrosurgical Unit (ESU). The system is intended for use by surgeons for the coagulation of cardiac and soft tissues during open surgical procedure.

e. Intended Use

The intended use for the Cobra Surgical Probes is as follows:

The Probe (Probe) is intended for the coagulation of cardiac tissue using radiofrequency (RF) energy during cardiac surgery. The Probe can be used during general surgery to coagulate soft tissues. The Probe may also be used to coagulate blood and soft tissue to produce hemostasis.

f. Comparison to Predicate Device

See Table I- Comparison of Device Characteristics to Predicate on the following page.

Table 1 - Comparison of Device Characteristics to Predicate

	Cobra® Surgical Probe(s)	Cobra® Surgical Probe(s)
Device Manufacturer & Name	Boston Scientific Corporation	Same
Device Description	Electrosurgical Probe	Same
510(k) Number	K013873	TBD
Regulatory Class	II	Same
Device Classification	21 CFR §878.4400	Same
Intended Use	Coagulation of Cardiac Tissue during Cardiac surgery and Soft Tissue during General Open Surgical Procedures	Same
Mode(s)	Monopolar	Same
Single Use?	Yes	Same
EO Sterilized?	Yes	Same
Shaft Size	8F	Same
Tip Material	Stainless Steel	Same
Length	15cm-35cm	Same
Electrode Size	6mm to 12.5 mm	Same
Electrode Number	2 to 7	Same

g. Summary of the Non-clinical Data

Where appropriate, testing conformed to the requirements of 21 CFR Part 58 (Good Laboratory Practices (GLP)). Specifically, non-clinical tests conducted for the Device showed the device met its design-input criteria, and was safe and effective for its intended use.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB 21 2008

Boston Scientific Corporation
c/o Ms. April M. Malmborg
Senior Specialist, Regulatory Affairs
2710 Orchard Parkway
San Jose, CA 95134

Re: K040219
Trade/Device Name: Cobra Surgical Probe
Regulation Number: 21 CFR 878.4400
Regulation Name: Electrosurgical cutting and coagulation device and accessories
Regulatory Class: II (two)
Product Code: OCL, GEI
Dated: January 30, 2004
Received: February 2, 2004

Dear Ms. Malmborg:

This letter corrects our substantially equivalent letter of February 24, 2004.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to continue marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html

Sincerely yours,



Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4. Premarket Notification -Indication for Use Statement

K040219

Device Name: Cobra Surgical Probes

Indication for Use:

The intended use for Cobra Surgical Probes is as follows:

The Probe is intended for the coagulation of cardiac tissue using radiofrequency (RF) energy during cardiac surgery. The Probe can be used during general surgery to coagulate soft tissues. The Probe may also be used to coagulate blood and soft tissue to produce hemostasis.

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IS NEEDED

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓ OR Over-the-Counter Use _____
(Per 21 CFR §801.109)

(Division Sign-Off)
Division of General, Restorative, and Neurological Devices

510(k) Number K040219